

Declaration of conformity

The manufacturer: **Score B.V.**
Feithspark 5
8356 BX Tolbert
The Netherlands
SRN **NL-MF-000006060**



declares under exclusive responsibility that the product

Product name: **Vena® Vario Levant**
Phlebotomy Chair - Bloedafnamestoel

Basic UDI-DI: 871722818VENAVLEVANT-01BM

Description: The Vena® Vario Levant Phlebotomy Chair is a medical device intended to support the body and to position the arm of the patient during phlebotomy, blood drawing, blood donation, plasma donation, or injection.
The Vena Vario Levant is a reclining phlebotomy chair with two phlebotomy armrests and Trendelenburg position. The Vena Vario Levant is a medical device that is adjustable in height medical device with configurable options and/or accessories.

Article number(s): 98xx.xxxxx. (xx.xxxxx referring to a configured article number)

has been designed, produced and tested according to the following harmonized standards for medical devices

ISO 9001:2015 Quality management systems
ISO 14791:2019 Medical devices - Application of risk management to medical devices

and is in compliance with all essential requirements of **Medical Device Regulation (MDR) EU 2017/745**

In accordance to the classification rules set out in this regulation the product is: **Class I**

Date: September 19, 2023
Place of issue: Tolbert, The Netherlands
Name: Mr. Mark Nauta
Title: Managing Director
Signature:

